

Replacing the Epidural Insert

Remove the camera.

Remove the simulator from the tray.

Push the old insert down and toward the front of the model.

Gently pull the old insert out.



Insert the new insert at the top first.

Gently push the insert into place using the flat of your hand.

Avoid using your fingers as you may damage the surface of the insert.